

Table 2.2.2-4. - Reference solutions Y

Reference solution	Volumes in millilitres	
	Standard solution Y	Hydrochloric acid (10 g/l HCl)
Y ₁	100.0	0.0
Y ₂	75.0	25.0
Y ₃	50.0	50.0
Y ₄	25.0	75.0
Y ₅	12.5	87.5
Y ₆	5.0	95.0
Y ₇	2.5	97.5

Table 2.2.2-5. - Reference solutions GY

Reference solution	Volumes in millilitres	
	Standard solution GY	Hydrochloric acid (10 g/l HCl)
GY ₁	25.0	75.0
GY ₂	15.0	85.0
GY ₃	8.5	91.5
GY ₄	5.0	95.0
GY ₅	3.0	97.0
GY ₆	1.5	98.5
GY ₇	0.75	99.25

Table 2.2.2-6. - Reference solutions R

Reference solution	Volumes in millilitres	
	Standard solution R	Hydrochloric acid (10 g/l HCl)
R ₁	100.0	0.0
R ₂	75.0	25.0
R ₃	50.0	50.0
R ₄	37.5	62.5
R ₅	25.0	75.0
R ₆	12.5	87.5
R ₇	5.0	95.0

Storage

For Method I, the reference solutions may be stored in sealed tubes of colourless, transparent, neutral glass of 12 mm external diameter, protected from light.

For Method II, prepare the reference solutions immediately before use from the standard solutions.

01/2005:20203

2.2.3. POTENTIOMETRIC DETERMINATION OF pH

The pH is a number which represents conventionally the hydrogen ion concentration of an aqueous solution. For practical purposes, its definition is an experimental one. The pH of a solution to be examined is related to that of a reference solution (pH_s) by the following equation:

$$\text{pH} = \text{pH}_s - \frac{E - E_s}{k}$$

in which E is the potential, expressed in volts, of the cell containing the solution to be examined and E_s is the potential, expressed in volts, of the cell containing the solution of known pH (pH_s), k is the change in potential per unit change in pH expressed in volts, and calculated from the Nernst equation.

Table 2.2.3-1. - Values of k at different temperatures

Temperature (°C)	k (V)
15	0.0572
20	0.0582
25	0.0592
30	0.0601
35	0.0611

The potentiometric determination of pH is made by measuring the potential difference between 2 appropriate electrodes immersed in the solution to be examined: 1 of these electrodes is sensitive to hydrogen ions (usually a glass electrode) and the other is the reference electrode (for example, a saturated calomel electrode).

Apparatus. The measuring apparatus is a voltmeter with an input resistance at least 100 times that of the electrodes used. It is normally graduated in pH units and has a sensitivity such that discrimination of at least 0.05 pH unit or at least 0.003 V may be achieved.

Method. Unless otherwise prescribed in the monograph, all measurements are made at the same temperature (20-25 °C). Table 2.2.3-2 shows the variation of pH with respect to temperature of a number of reference buffer solutions used for calibration. For the temperature correction, when necessary, follow the manufacturer's instructions. The apparatus is calibrated with the buffer solution of potassium hydrogen phthalate (primary standard) and 1 other buffer solution of different pH (preferably one shown in Table 2.2.3-2). The pH of a third buffer solution of intermediate pH read off on the scale must not differ by more than 0.05 pH unit from the value corresponding to this solution. Immerse the electrodes in the solution to be examined and take the reading in the same conditions as for the buffer solutions.

When the apparatus is in frequent use, checks must be carried out regularly. If not, such checks should be carried out before each measurement.

All solutions to be examined and the reference buffer solutions must be prepared using *carbon dioxide-free water R*.

PREPARATION OF REFERENCE BUFFER SOLUTIONS

Potassium tetraoxalate 0.05 M. Dissolve 12.61 g of C₄H₃KO₈·2H₂O in *carbon dioxide-free water R* and dilute to 1000.0 ml with the same solvent.

Potassium hydrogen tartrate, saturated at 25 °C. Shake an excess of C₄H₅KO₆ vigorously with *carbon dioxide-free water R* at 25 °C. Filter or decant. Prepare immediately before use.

Potassium dihydrogen citrate 0.05 M. Dissolve 11.41 g of C₆H₇KO₇ in *carbon dioxide-free water R* and dilute to 1000.0 ml with the same solvent. Prepare immediately before use.

Potassium hydrogen phthalate 0.05 M. Dissolve 10.13 g of C₈H₅KO₄, previously dried for 1 h at 110 ± 2 °C, in *carbon dioxide-free water R* and dilute to 1000.0 ml with the same solvent.

Table 2.2.3.2. – pH of reference buffer solutions at various temperatures

Temperature (°C)	Potassium tetraoxalate 0.05 M	Potassium hydrogen tartrate saturated at 25 °C	Potassium dihydrogen citrate 0.05 M	Potassium hydrogen phthalate 0.05 M	Potassium dihydrogen phosphate 0.025 M + disodium hydrogen phosphate 0.025 M	Potassium dihydrogen phosphate 0.0087 M + disodium hydrogen phosphate 0.0303 M	Disodium tetraborate 0.01 M	Sodium carbonate 0.025 M + sodium bicarbonate 0.025 M	Calcium hydroxide, saturated at 25 °C
	$C_4H_3KO_8 \cdot 2H_2O$	$C_4H_5KO_6$	$C_6H_7KO_7$	$C_8H_5KO_4$	$KH_2PO_4 + Na_2HPO_4$	$KH_2PO_4 + Na_2HPO_4$	$Na_2B_4O_7 \cdot 10H_2O$	$Na_2CO_3 + NaHCO_3$	$Ca(OH)_2$
15	1.67		3.80	4.00	6.90	7.45	9.28	10.12	12.81
20	1.68		3.79	4.00	6.88	7.43	9.23	10.06	12.63
25	1.68	3.56	3.78	4.01	6.87	7.41	9.18	10.01	12.45
30	1.68	3.55	3.77	4.02	6.85	7.40	9.14	9.97	12.29
35	1.69	3.55	3.76	4.02	6.84	7.39	9.10	9.93	12.13
$\frac{\Delta pH^{(1)}}{\Delta t}$	+ 0.001	- 0.0014	- 0.0022	+ 0.0012	- 0.0028	- 0.0028	- 0.0082	- 0.0096	- 0.034

(1) pH variation per degree Celsius.

Potassium dihydrogen phosphate 0.025 M + disodium hydrogen phosphate 0.025 M. Dissolve 3.39 g of KH_2PO_4 and 3.53 g of Na_2HPO_4 , both previously dried for 2 h at 120 ± 2 °C, in carbon dioxide-free water R and dilute to 1000.0 ml with the same solvent.

Potassium dihydrogen phosphate 0.0087 M + disodium hydrogen phosphate 0.0303 M. Dissolve 1.18 g of KH_2PO_4 and 4.30 g of Na_2HPO_4 , both previously dried for 2 h at 120 ± 2 °C, in carbon dioxide-free water R and dilute to 1000.0 ml with the same solvent.

Disodium tetraborate 0.01 M. Dissolve 3.80 g of $Na_2B_4O_7 \cdot 10H_2O$ in carbon dioxide-free water R and dilute to 1000.0 ml with the same solvent. Store protected from atmospheric carbon dioxide.

Sodium carbonate 0.025 M + sodium hydrogen carbonate 0.025 M. Dissolve 2.64 g of Na_2CO_3 and 2.09 g of $NaHCO_3$ in carbon dioxide-free water R and dilute to 1000.0 ml with the same solvent. Store protected from atmospheric carbon dioxide.

Calcium hydroxide, saturated at 25 °C. Shake an excess of calcium hydroxide R with carbon dioxide-free water R and decant at 25 °C. Store protected from atmospheric carbon dioxide.

STORAGE

Store buffer solutions in suitable chemically resistant, tight containers, such as type I glass bottles or plastic containers suitable for aqueous solutions.

Table 2.2.4.1

Reaction	pH	Indicator	Colour
Alkaline	> 8	Litmus paper red R	Blue
		Thymol blue solution R (0.05 ml)	Grey or violet-blue
Slightly alkaline	8.0 – 10.0	Phenolphthalein solution R (0.05 ml)	Colourless or pink
		Thymol blue solution R (0.05 ml)	Grey
Strongly alkaline	> 10	Phenolphthalein paper R	Red
		Thymol blue solution R (0.05 ml)	Violet-blue
Neutral	6.0 – 8.0	Methyl red solution R	Yellow
		Phenol red solution R (0.05 ml)	
Neutral to methyl red	4.5 – 6.0	Methyl red solution R	Orange-red
Neutral to phenolphthalein	< 8.0	Phenolphthalein solution R (0.05 ml)	Colourless; pink or red after adding 0.05 ml of 0.1 M base
Acid	< 6	Methyl red solution R	Orange or red
		Bromothymol blue solution RI	Yellow
Faintly acid	4.0 – 6.0	Methyl red solution R	Orange
		Bromocresol green solution R	Green or blue
Strongly acid	< 4	Congo red paper R	Green or blue

01/2005:20204

01/2005:20205

2.2.4. RELATIONSHIP BETWEEN REACTION OF SOLUTION, APPROXIMATE pH AND COLOUR OF CERTAIN INDICATORS

To 10 ml of the solution to be examined, add 0.1 ml of the indicator solution, unless otherwise prescribed in Table 2.2.4.-1.

2.2.5. RELATIVE DENSITY

The relative density $d_{t_2}^{t_1}$ of a substance is the ratio of the mass of a certain volume of a substance at temperature t_1 to the mass of an equal volume of water at temperature t_2 . Unless otherwise indicated, the relative density d_{20}^{20} is used. Relative density is also commonly expressed as d_4^{20} . Density ρ_{20} , defined as the mass of a unit volume of the substance at 20 °C may also be used, expressed in kilograms per cubic metre or grams per cubic centimetre