

Figure 2.9.1.-2. – Disintegration apparatus B  
Dimensions in millimetres

01/2008:20902

## 2.9.2. DISINTEGRATION OF SUPPOSITORIES AND PESSARIES

The disintegration test determines whether the suppositories or pessaries soften or disintegrate within the prescribed time when placed in a liquid medium in the experimental conditions described below.

Disintegration is considered to be achieved when:

- a) dissolution is complete,
- b) the components of the suppository or pessary have separated: melted fatty substances collect on the surface of the liquid, insoluble powders fall to the bottom and soluble components dissolve, depending on the type of preparation, the components may be distributed in one or more of these ways,

c) there is softening of the sample that may be accompanied by appreciable change of shape without complete separation of the components, the softening is such that the suppository or pessary no longer has a solid core offering resistance to pressure of a glass rod,

d) rupture of the gelatin shell of rectal or vaginal capsules occurs allowing release of the contents,

e) no residue remains on the perforated disc or if a residue remains, it consists only of a soft or frothy mass having no solid core offering resistance to pressure of a glass rod (vaginal tablets).

**Apparatus.** The apparatus (Figure 2.9.2.-1) consists of a sleeve of glass or suitable transparent plastic, of appropriate thickness, to the interior of which is attached by means of three hooks a metal device consisting of two perforated stainless metal discs each containing 39 holes 4 mm in diameter; the diameter of the discs is similar to that of the interior of the sleeve; the discs are about 30 mm apart. The test is carried out using three such apparatuses each containing a single sample. Each apparatus is placed in a

**Method.** Use three suppositories or pessaries. Place each one on the lower disc of a device, place the latter in the sleeve and secure. Invert the apparatuses every 10 min. Examine the samples after the period prescribed in the monograph. To pass the test all the samples must have disintegrated.





Figure 2.9.2-1. – Apparatus for disintegration of  
suppositories and pessaries  
Dimensions in millimetres

Use the apparatus described above, arranged so as to rest on the hooks (see Figure 2.9.2-2). Place it in a beaker of suitable diameter containing water maintained at 36-37 °C with the level just below the upper perforated disc. Using a pipette, adjust the level with water at 36-37 °C until a uniform film covers the perforations of the disc. Use three vaginal tablets. Place each one on the upper plate of an apparatus and cover the latter with a glass plate to maintain appropriate conditions of humidity. Examine the state of the samples after the period prescribed in the monograph. To pass the test all the samples must have disintegrated.



- A. glass plate  
B. vaginal tablet  
C. water surface  
D. water  
E. dish, beaker

Figure 2.9.2.-2.

**01/2008:20903**

### 2.9.3. DISSOLUTION TEST FOR SOLID DOSAGE FORMS

This test is provided to determine compliance with the dissolution requirements for solid dosage forms administered orally. In this chapter, a dosage unit is defined as 1 tablet or 1 capsule or the amount specified.

## APPARATUS

**Apparatus 1 (Basket apparatus).** The assembly consists of the following: a vessel, which may be covered, made of glass or other inert, transparent material<sup>(1)</sup>; a motor; a drive shaft; and a cylindrical basket (stirring element). The vessel is partially immersed in a suitable water-bath of any convenient size or heated by a suitable device such as a heating jacket. The water-bath or heating device permits maintaining the temperature inside the vessel at  $37 \pm 0.5^\circ\text{C}$  during the test and keeping the dissolution medium in constant, smooth motion. No part of the assembly, including the environment in which the assembly is placed, contributes significant motion, agitation, or vibration beyond that due to the smoothly rotating stirring element. Apparatus that permits observation of the preparation and stirring element during the test is preferable. The vessel is cylindrical, with a hemispherical bottom and a capacity of 1 litre. Its height is 160-210 mm and its inside diameter is 98-106 mm. Its sides are flanged at the top. A fitted cover may be used to retard evaporation<sup>(2)</sup>. The shaft is positioned so that its axis is not more than 2 mm at any point from the vertical axis of the vessel and rotates smoothly and without significant wobble that could affect the results. A speed-regulating device is used that allows the shaft rotation speed to be selected and maintained at a specified rate, within  $\pm 4$  per cent.

Shaft and basket components of the stirring element are fabricated of stainless steel, type 316 or equivalent, to the specifications shown in Figure 2.9.3.-1.

(1) The materials must not sorb, react, or interfere with the preparation to be tested.

- (1) The materials must not sorb, react, or interfere with the preparation to be tested.
- (2) If a cover is used, it provides sufficient openings to allow ready insertion of the thermometer and withdrawal of samples.