CRITERIA

Apply the following criteria, unless otherwise specified.

Solid and liquid dosage forms. The requirements for dosage uniformity are met if the acceptance value of the first 10 dosage units is less than or equal to L1. If the acceptance value is greater than L1, test the next 20 dosage units and calculate the acceptance value. The requirements are met if the final acceptance value of the 30 dosage units is less than or equal to L1 and no individual content of the dosage unit is less than $(1 - L2 \times 0.01)M$ nor more than $(1 + L2 \times 0.01)M$ in calculation of acceptance value under content uniformity or under mass variation. Unless otherwise specified, L1 is 15.0 and L2 is 25.0.

01/2008:20941

2.9.41. FRIABILITY OF GRANULES AND SPHEROIDS

This chapter describes 2 methods for determination of the friability of granules and spheroids, which may be used during development studies. It is recognised, however, that many methods with equal suitability may be used.

This test is intended to determine, under defined conditions, the friability of granules and spheroids. Friability is defined as a reduction in the mass of the granules or spheroids or in the formation of fragments of granules or spheroids, occurring when the granules or spheroids are subjected to mechanical strain during handling (tumbling, vibration, fluidisation, etc.). Examples of changes are abrasion, breakage or deformation of granules or spheroids.

METHOD A

Apparatus (fluidised-bed apparatus). The apparatus (see Figure 2.9.41.-1) consists of a glass cylinder (A) with a conical lower part. The cylinder is provided with a sieve lid (B) having an aperture size of 500 μ m or any other suitable sieve. The conical end is connected to a U-shaped glass tube (C) that can be disconnected from the cylinder for removal of the granules or spheroids. The U-tube is attached to a T-coupling (D). One inlet of the T-coupling is joined by a silicone tube to a manometer for regulating the compressed-air flow (use compressed air complying with the test for water in the monograph *Medicinal air (1238)*), the other one is connected via a silicone tube to a by-pass flowmeter (E) (0.10-1.00 m³h⁻¹).

Procedure. The following procedure is usually suitable. Remove the fine particles by sieving (sieve having an aperture size of 710 µm or any other suitable sieve). Introduce about 8.0 g (m_1) of granules or spheroids into the cylinder (A). Close the apparatus with the sieve lid (B). Adjust the flow rate of the compressed air to 0.45 m³h⁻¹. After 15 min, remove the granules or spheroids from the apparatus by disconnecting the U-tube and weigh again (m_2) . Test 3 samples and calculate the mean value. It is recommended




Figure 2.9.41.-1. - Fluidised-bed apparatus

to spray the inside of the apparatus with an antistatic agent every 3 determinations in order to prevent electrostatic charging.

Loss on drying. Dry in an oven at 105 °C, unless otherwise prescribed. Alternatively, other drying conditions as described in general method *2.2.32* may be used.

Calculation

$$F = \frac{m_1 (100 - T_1) - m_2 (100 - T_2)}{m_1} \times 100$$

F = friability;

- T_1 = percentage loss on drying before the test (mean of 2 determinations);
- T_2 = percentage loss on drying after the test (mean of 2 determinations);
- m_1 = mass of the granules or spheroids before the test, in grams;
- m_2 = mass of the granules or spheroids after the test, in grams.

METHOD B

Apparatus (oscillating apparatus). The apparatus (see Figure 2.9.41.-2) consists of a 105 ml glass container, containing the granules or spheroids to be examined, which is subjected to horizontal oscillations. The frequency and duration of the oscillations can be varied continuously. The

frequency can be adjusted, using a scale, to a value in the range 0-400 oscillations/min. The duration can be set to a value in the range 0-9999 s.

Procedure. The following procedure is usually suitable. Remove the fine particles by sieving (sieve having an aperture size of 355 μ m or any other suitable sieve). In the glass container, weigh about 10.00 g (m_1) of the granules or spheroids. Install the container in the apparatus. Shake for 240 s at the highest frequency for hard granules or spheroids, or for 120 s at a lower frequency (e.g. 140 oscillations/min) for soft granules or spheroids. Sieve (355 μ m, or the same sieve as used previously) and weigh the granules or spheroids again (m_2). Test 3 samples and calculate the mean value.

Loss on drying. Dry in an oven at 105 $^{\circ}$ C, unless otherwise prescribed. Alternatively, other drying conditions as described in general method *2.2.32* may be used.

Calculation

$$F = \frac{m_1 \left(100 - T_1\right) - m_2 \left(100 - T_2\right)}{m_1} \times 100$$

- F = friability;
- T_1 = percentage loss on drying before the test (mean of 2 determinations);
- T_2 = percentage loss on drying after the test (mean of 2 determinations);
- m_1 = mass of the granules or spheroids before the test, in grams;
- $m_2 = \max_{i=1}^{n} m_{i}$ mass of the granules or spheroids after the test, in grams.




Figure 2.9.41.-2. - Oscillating apparatus